

## TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage two Envelopes
3. Title of Procurement: Procurement of Laboratory Equipment for Haematology Department, University of Health Sciences, Lahore
4. Tender Inquiry No: UHS/Pur/1062/2021
5. Date & Time of Bid Closing: 17.02.2021 at 11:00 AM
6. Date & Time of Bid Opening 17.02.2021 at 11:30 AM
7. Nos. of Bids Received: 04

**TECHNICAL EVALUATION REPORT (HAEMATOLOGY DEPARTMENT)**

Sr#	Items	Specifications	Vendors			
			S. Ejazuddin & Co	Reaction Scientific (Private) Limited	Nex Tek Health Care	RAAB Scientific Solutions
01	Fully Automated Haematology Analyzer for CBC	<p><b><u>Detailed Requirements:</u></b></p> <ul style="list-style-type: none"> <li>Automated Hematology Counter</li> <li>Auto probe cleaning/wiping</li> <li>RS232/USB interface</li> <li>220V, 50 Hz operated</li> </ul> <p><b><u>User Adjustable Settings:</u></b></p> <ul style="list-style-type: none"> <li>Through put 100 samples/hour or more</li> <li>Minimum 24 reportable parameters 5-Part differential including Hb, WBC, RBC, Hct, PLT, derived values and must include immature granulocytes, absolute NRBCs, absolute</li> </ul>	Quoted as per BOQs <i>Qualified</i>	Quoted as per BOQs <i>Qualified</i>	Quoted as per BOQs <i>Qualified</i>	Not quoted

**TECHNICAL EVALUATION REPORT (HAEMATOLOGY DEPARTMENT)**

		<p>Reticulocytes, IPF (or equivalent) and IRF (or equivalent)</p> <ul style="list-style-type: none"><li>• Histograms for WBC, RBC and Platelet</li><li>• Sample Volume: Max 120 ul of whole blood or less</li></ul> <p><b><u>Displayed Parameters:</u></b></p> <ul style="list-style-type: none"><li>• LCD display with touch screen</li><li>• Patient data entry and specimen recognition by keyboard/LCD Touch Screen/Bar Code</li><li>• Patient data and results storage for 10,000 Results or more</li></ul> <p><b><u>Accessories:</u></b></p> <ul style="list-style-type: none"><li>• Autoloader sampler, capacity at least 50 samples</li><li>• Separate specimen mixer</li><li>• Separate printer required</li><li>• Separate computer required</li></ul>				
--	--	--	--	--	--	--