


TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage two Envelopes
3. Title of Procurement: Procurement of Laboratory Research Equipment of Pharmacology Department, University of Health Sciences, Lahore
4. Tender Inquiry No: UHS/Pur/930/2020
5. Date & Time of Bid Closing: 15-06-2020 at 11:00 AM
6. Date & Time of Bid Opening: 15-06-2020 at 11:30 AM
7. Nos. of Bids Received: 01

Sr. No	Item	Specifications	Vendor participated
01	Desiccator with plates	Nalgene 246 mm, 4/case, cabinet type, 18x12x12" Europe/UK/USA origin/Equivalent	Quoted as per BOQ Qualified
02	Drying Board	 <p>Wall Desk Drying Rack Number of pegs per rack: 27 Pegs. Features of pegs: detachable drying pegs, dimensions: 5inch (125mm) long, 0.35inch (9mm) dia. 6 rows of pegs, staggered rows. Built-in molded draining tray 22inch (550mm) x4.8 inch (120mm), with hole in the tray for drainage, supplied with 30 inch (760mm) of vinyl tubing. Rack dimensions: 22 inch (550mm) high x 16 inch (400mm) wide. Supplied complete with wall mounting bracket. Europe/UK/USA origin/Equivalent</p>	Quoted as per BOQ Qualified
03	PCR work Station	<p>External dimensions (WxDxH): 32x24x39 inch Internal dimensions (WxDxH): 31x23x27 inch</p> <ul style="list-style-type: none"> • UVTect™ microprocessor controller • No installation required - plugs directly into a standard 110V or 220V electrical outlet • ISO 5 vertical laminar flow air • No ductwork required • UV shelf with built-in pipettor holder • Seamless plastic design -no joints or gaps in construction • Lab event timer • Available in standard height or tall version • Clear polycarbonate shell for 360° visibility • Extremely quiet operation < 49dBA • UV and fluorescent lights • Digital UV light timer :01-59 minutes • Access port for electrical cords <p>Europe/UK/USA origin/Equivalent</p>	Quoted as per BOQ Qualified

All technical bids were evaluated against the laid specifications and comparative statement (summary/detail) is being submitted to Chairman Tender Committee and Incharge Purchase Cell for further necessary action.

No Complaint Shall be entertained after 24 hours of display of Technical Evaluation report.

*For Dr. Shahzad
Chaudhry*



Dr. Allah Rakha
Chairman, Tender Committee

Prof Dr. M. Shahzad
HOD Pharmacology

For Genu
Prof. Dr. Shagufta Khaliq
Incharge, Purchase Cell