

TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage two Envelopes
3. Title of Procurement: Procurement of Laboratory Research Equipment for ELISA based Covid 19 tests by University of Health Sciences, Lahore
4. Tender Inquiry No: UHS/Pur/962/2019
5. Date & Time of Bid Closing: 18-09-2020 at 11:00 AM
6. Date & Time of Bid Opening: 18-09-2020 at 11:30 AM
7. Nos. of Bids Received: 09

S#	Item	Pakistan Hospital	Mod Scientific	Bio-Sphere	Rays Technology	Sameer Enterprises	Worldwide Scientific	H.A. Shah & Sons	Sure Bio Diagnostics & Pharmaceuticals	Muslim Medical Services
1	Safety Cabinet Class-2 A-2	Not Quoted	Not Quoted	Not Qualified (Pre-Filter incoming Air not present & brochure is not enclosed)	Qualified if provide NSF49 or ISO USFDA or EN12469 certification	Qualified if provide NSF49 or ISO USFDA or EN12469 certification	Qualified if provide NSF49 or ISO USFDA or EN12469 certification	Not Quoted	Not Qualified Working area is less than required	Qualified
2	Autoclave	Not Quoted	Not Quoted	Not Qualified (Pressure range is less than required)	Qualified	Not Qualified (Temperature range & sterilization time is not according to specification)	Not Qualified (Pressure range and temperature is less than required)	Not Quoted	Not Qualified (Pressure range is less than required)	Not Quoted
3	Refrigerated Centrifuge	Not Quoted	Not Quoted	Not Quoted	Not Qualified 1- Built in condensate drain to eliminate water accumulation & prevent corrosion is not present 2- 5 ml adaptor quoted, pcr plate rotor quoted as optional(not included in the offer) 3- Noise level is heighr than required	Not Quoted	Not Qualified Rotors are not according to specifications	Qualified	Not Qualified 1- Built in condensate drain to eliminate water accumulation & prevent corrosion is not present 2- Noise level is heighr than required 3- Must have "At set rpm" function which enables timer countdown to be started only when selected speed is achieved is not present	Not Quoted
4	Refrigerated Centrifuge (For Eppendorf)	Not Qualified speed is less than required)	Not Qualified speed is less than required)	Not Qualified speed is less than required)	Not Quoted	Not Qualified speed is less than required)	Not Qualified speed with rotor is less than required	Not Qualified speed is less than required)	Not Qualified speed is less than required)	Not Quoted
5	Water Incubator with Shaking Tray	Not Quoted	Not Quoted	Not Qualified (Additional Boiling Mood in temperature range is not mentioned)	Qualified	Not Quoted	Not qualified Temperature range is not according to specs	Not Quoted	Not Quoted	Not Quoted
6	Refrigerator (2-8 C)	Not Quoted	Not Quoted	Not Quoted	Qualified if provide Battery back up	Not Quoted	Not Quoted	Not Quoted	Qualified if provide Battery back up	Qualified
7	Refrigerator (-20C)	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Not Qualified (Volume is less than required)	Not Quoted	Not Quoted	Qualified	Not Qualified (Volume is less than required)
8	Incubator	Not Quoted	Not Quoted	Qualified	Qualified	Not Quoted	Not Qualified internal diemensions are not according to specs	Not Quoted	Qualified	Not Quoted
9	Vortex Mixer	Not Quoted	Not Qualified (Speed is less than required)	Qualified	Not Qualified (Speed is less than required)	Not Qualified (Speed is less than required)	Not Qualified (Speed is less than required)	Not Quoted	Not Qualified (Speed is less than required)	Not Quoted
10	Dispensing Bottles	Not Quoted	Qualified	Qualified	Qualified	Qualified	Qualified	Not Quoted	Not Quoted	Not Quoted
11	UV Sterilizer Disinfection Cabinet	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Not according to Specifications	Quote PCR Station rather than required item	Not Quoted	Not Quoted	Not Quoted
12	Pipette set (P2, P10, P20, P200, P1000)	Qualified	Not Qualified (not equivalent to specifications because all ranges are not available))	Not Qualified (not equivalent to specifications because all ranges are not available))	Qualified	Not Qualified (not equivalent to specifications because all ranges are not available))	Qualified	Qualified	Qualified	Not Quoted
13	Pipette Tips (Yellow, Blue, White Tips)	Qualified	Specifications not mentioned	Not Quoted	Not Quoted	Qualified	Qualified	Not Quoted	Not Quoted	Not Quoted
14	Centrifuge Tube racks (Polypropylene)	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Qualified	Not Quoted	Not Quoted	Qualified	Not Quoted

Asif

All technical bids were evaluated against the laid specifications and comparative statement (summary/detail) is being submitted to Chairman Tender Committee and Incharge Purchase Cell for further necessary action.

No Complaint Shall be entertained after 07 days of display of Technical Evaluation report.



Prof. Dr. Nadem Afzal
HOD Immunology Department



Dr. Allah Rakha
Chairman, Tender Committee



Prof. Dr. Shagufta Khaliq
Incharge Purchase Cell